



中国认可
国际互认
校准
CALIBRATION
CNAS L0599

证书编号:
Certificate No.



820419409A

通标标准技术服务(上海)有限公司

SGS-CSTC Standards Technical Services (Shanghai) Co., Ltd.

校准证书

Calibration Certificate

委 托 者
Customer

笛尔流体科技(衡水)有限公司
DME AND JDE FLUID CO.,LTD

仪 器 名 称
Equipment

数字压力计
Digital pressure gauge

型 号 / 规 格
Model/Specification

MD-S211

机 身 编 号
Serial Number

DE770F024AB0

管 理 编 号
Management Number

/

制 造 厂 商
Manufacturer

XZT

委托者地址
Address of customer

河北省衡水市深州市前么头镇前么头村西北角(地方铁路以北)
Qian Mo Tou Village, Qian Mo Tou Town, Shenzhou, Hengshui City, Hebei

批 准
Approved by

人

杨峻锋

杨峻锋

核 验
Checked by

员

黄红霞

黄红霞

校 准
Calibrated by

员

武云海

武云海

接收日期
Received Date

2025

年
Year

09

月
Month

30

日
Day

校准日期
Calibration Date

2025

年
Year

10

月
Month

10

日
Day

发布日期
Issued Date

2025

年
Year

10

月
Month

10

日
Day

Unless otherwise stated the results shown in this calibration certificate refer only to the sample(s) calibrated.

Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <https://www.sgs.com/en/Terms-and-Conditions>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com





■ 声明

Statement

- 证书中的校准结果均可溯源至国际单位制(SI)单位和社会公用计量标准。
All the calibration results in this certificate can be traced to the international system of units(SI)units and/or social public meterage standards.
- 本校准结果, 仅对受校样品的本次校准有效。
It's effect that the results of this report relate only to the sample(s) calibrated.
- 未经SGS批准, 部分采用本证书内容无效。
Partly using this calibration certificate will not be admitted unless allowed by SGS.
- ***项目为本公司校准实验室具备的能力, 而非中国合格评定国家认可委员会的认可能力。
***project representative calibration laboratory in this company have the capacity, rather than CNAS accreditation capacity.
- 证书页中, "P"代表"符合", "F"代表"不符合", "N/A"代表"不适用", "****E*"代表"****×10**".
In the data sheet, "P"represents"Pass", "F"represents"Fail", "N/A"represents"Not Applicable", "****E*"represents "****×10**".
- 除非另有说明, 使用简单接受的二元判定规则进行符合性声明。
Unless otherwise stated, the decision rule for conformity reporting is based on Binary Statement for Simple Acceptance Rule.

■ 本次校准的技术依据

Reference documents

JJG 875-2019《数字压力计检定规程》

JJG 875-2019《Verification Regulation of Digital Pressure Gauges》

■ 本次校准所使用的主要测量标准

Main measurement standards used during the calibration

仪器名称/型号
Equipment/ Model编号/有效期至
Equipment No./ Due Date测量范围/技术指标
Measuring range/ Technical Specifications溯源机构/证书编号
Verification or Certification Organization/Certificate No.智能数字压力表
ConST211ASHCAL-1734 (211A2500
2440)
2026-05-20

(0~250) MPa 0.05级

北京恒矩检测技术有限公司
C202505210027

■ 校准地点/Calibration Location: 上海市浦东新区康桥东路1159弄69号10号楼 压力室

■ 环境条件/Environmental condition during the calibration:

温度(Temperature): (21.5~21.7)℃

湿度(Humidity): (52~55)%RH

其他(other): /

■ 备注/Comments:

本证书或报告替代820419409编号证书或报告, 原820419409编号证书或报告作废, 本次修改时间为2025年10月15日。
This certificate/report supersedes the certificate/report No.820419409, original certificate/report will be invalid from date 2025-10-15.

Unless otherwise stated the results shown in this calibration certificate refer only to the sample(s) calibrated.

Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <https://www.sgs.com/en/Terms-and-Conditions>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.

Attention: To check the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com





校准结果(Calibration results):

外观检查/Appearance Inspection: 符合要求/Meet Requirements.

示值误差/Indication Error:

等级/Class: 1

标准值 Standard Value (bar)	示值误差 Indication Error (bar)	回程误差 Return Error (bar)	扩展不确定度 Uncertainty $U(k=2)(\%FS)$	允许误差 Limit (bar)	结论 Conclusion
0	0	0	0.2	± 20	P
500	-1	0	0.2	± 20	P
1000	-2	1	0.2	± 20	P
1500	-2	0	0.2	± 20	P
2000	-4	1	0.2	± 20	P

说明/Remark:

1bar=0.1MPa

本次校准结果符合上述技术要求。

The calibration results accord with above requirements.

建议复校周期: 12个月。

Recommend recalibration period :12 months.

---校准结果内容结束---

---THE END---

Unless otherwise stated the results shown in this calibration certificate refer only to the sample(s) calibrated.

Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <https://www.sgs.com/en/Terms-and-Conditions>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.

Attention: To check the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

